

Program MicroLog 3/5.2.01 35
 Project ML5
 File Name Vojtkova RPP 2021-2.D5E
 User Vojtkova
 Instrument MicroStation 2 Reader
 Instrument S/N 18032122
 Incubation Hours 25.00
 Plate Number 2
 Plate Type GEN III
 Protocol A

Cislo vzorku 6A3
 Datum 01.06.2021
 Lokalita Ostrava DP Ecocoal
 Cas pripravy 31.05.2021 12.10
 Zpracoval Vojtková

Date & Time of Read Jun 01 2021 12:23 PM
 Biolog ID DB GEN_III_v2.8.0.I5G

Result Species ID: Pseudomonas chlororaphis ss aurantiaca
 Comment
 Notice

Rank	PROB	SIM	DIST	Organism Type	Species
1	0.907	0.653	5.051	GN-Nent	Pseudomonas chlororaphis ss aurantiaca
2	0.070	0.191	5.475	GN-Nent	Pseudomonas marginalis
3	0.014	0.085	6.182	GN-Nent	Pseudomonas putida
4	0.010	0.071	6.336	GN-Nent	Pseudomonas corrugata

Key: <x: positive, x: negative, <x-: mismatched positive, x+: mismatched negative, {x: borderline, -x: less than A1 well

Well Color Values

Plate	1	2	3	4	5	6	7	8	9	10	11	12
A	61	80	54	{ 150	53	60	76	58	52	< 287	< 267	33 +
B	56	48	71	58	54	{ 104 +	49	50	57	< 274	72	40
C	< 199	{ 158	< 163	{ 126	83	{ 124	{ 118	79	{ 94	< 304	{ 188	< 263
D	60	{ 151	65	< 170	{ 141	74	{ 127	55	< 222	< 256	< 268	{ 209
E	61	{ 94	< 212	{ 151	< 195	< 212	< 213	{ 96 +	< 184	< 281	{ 156	{ 224
F	57	{ 111	{ 103	< 211	{ 111	< 162 -	< 197	< 246	< 191	< 283	< 396	< 393
G	< 186	{ 115	68	< 189	< 218	{ 120	67	< 197	{ 95	< 273	{ 148	< 252
H	{ 144	< 172	64	< 170	68	65	< 169	< 176	88	{ 221 -	45	39

Report Date Jun 01 2021 12:23 PM